

Table 2: Procedure characteristics and parameters at implantation of the leadless pacemakers.

| | |
|----------------------------------|-----------------|
| Procedure time (min) | 66 ± 12 |
| Fluoroscopy time (min) | 9 [7-12] |
| Number of device deployments (%) | |
| 1 | 5 (56 %) |
| 2 | 3 (33 %) |
| 3 | 1 (11 %) |
| Sensing amplitude (mV) | 11.4 ± 6.8 |
| Pacing threshold (V at 0.24ms) | 0.43 ± 0.16 |
| Pacing impedance (Ω) | 789 ± 118 |