

Table 1: Intra-day precision results regarding the chemiluminescence analyzer performance based on the Clinical and Laboratory Standards Institute EP15-A2 document requirements.

Test items	Low concentration			High concentration		
	\bar{x}	s	$CV/\%$	\bar{x}	s	$CV/\%$
TM (TU/mL)	20.44	0.37	1.79	78.83	1.01	1.28
TAT (ng/mL)	10.28	0.18	1.76	40.40	0.36	0.88
PIC ($\mu\text{g/mL}$)	2.01	0.05	2.48	7.33	0.17	2.31
t-PAIC (ng/mL)	4.79	0.10	2.07	18.09	0.28	1.55

TM: thrombomodulin, TAT: thrombin-antithrombin complex, PIC: plasmin- α 2-plasmin inhibitor complex, tPAIC: tissue plasminogen activator/plasminogen activator inhibitor-1, CV: coefficient of variation, \bar{x} : mean, s : standard deviation.