

Table 2: Inter-day precision results regarding the chemiluminescence analyzer performance based on the Clinical and Laboratory Standards Institute EP15-A2 document requirements.

Test items	Low concentration			High concentration		
	\bar{x}	s	$CV/\%$	\bar{x}	s	$CV/\%$
TM (TU/mL)	20.38	0.55	2.68	78.28	1.61	2.06
TAT (ng/mL)	9.97	0.32	3.23	40.16	0.91	2.27
PIC ($\mu\text{g/mL}$)	1.98	0.07	3.69	7.16	0.18	2.53
t-PAIC (ng/mL)	4.81	0.07	1.41	18.32	0.32	1.73

TM: thrombomodulin, TAT: thrombin-antithrombin complex, PIC: plasmin- α 2-plasmin inhibitor complex, tPAIC: tissue plasminogen activator/plasminogen activator inhibitor-1, CV: coefficient of variation, \bar{x} : mean, s : standard deviation.