

Table 4: Reference interval result regarding the chemiluminescence analyzer performance based on the Clinical and Laboratory Standards Institute C28-A document requirements.

Test items	Minimum	Maximum	Reference range
TM (TU/mL)	4.5	10.6	3.8~13.3
TAT (ng/mL)	0.3	3.2	<4.0
PIC (μ g/mL)	0.32	0.71	<0.8
t-PAIC (ng/mL)	female 3.1 male 1.4	female 9.9 male 11.8	female <10.5 male <17.0

TM: thrombomodulin, TAT: thrombin-antithrombin complex, PIC: plasmin- α 2-plasmin inhibitor complex, tPAIC: tissue plasminogen activator/plasminogen activator inhibitor-1.